To avoid burns near heat, insert the stem into the pocket holder's built-in wrench and use the pocket holder as a handle. DO NOT rotate the head.

TO USE THE POCKET CASE AS A HOLDER

TO USE THE RE-CAL POCKET CASE FOR RECALIBRATION The thermometer will withstand drops to the

shift. If, however, it appears to be inaccurate, immerse the thermometer stem at least 2" in an ice slurry (32°F). Determine the degrees of inaccuracy. Then with the

pocket case is not handy, a wrench or pliers

may be used instead.

floor without damage or major calibration hex nut on the pocket case, rotate the thermometer head to correct the error. If the