To avoid burns near heat, insert the stem into the pocket holder's built-in wrench and use the pocket holder as a handle. DO NOT rotate the head.

TO USE THE RE-CAL POCKET CASE FOR RECALIBRATION The thermometer will withstand drops to the floor

TO USE THE POCKET CASE AS A HOLDER

without damage or major calibration shift. If, however, it appears to be inaccurate, immerse the thermometer stem at least 2" in an ice slurry (32°F).

Determine the degrees of inaccuracy. Then with the hex nut on the pocket case, rotate the thermometer head to correct the error. If the pocket case is not

handy, a wrench or pliers may be used instead.